



NBS TECHNICAL NOTE 856

U.S. DEPARTMENT OF COMMERCE / National Bureau of Standards

Note on a Vibratory Phenomenon Arising in Transducer Calibration

QC

100

45753

no. 856

1975

C.2

NATIONAL BUREAU OF STANDARDS

The National Bureau of Standards¹ was established by an act of Congress March 3, 1901. The Bureau's overall goal is to strengthen and advance the Nation's science and technology and facilitate their effective application for public benefit. To this end, the Bureau conducts research and provides: (1) a basis for the Nation's physical measurement system, (2) scientific and technological services for industry and government, (3) a technical basis for equity in trade, and (4) technical services to promote public safety. The Bureau consists of the Institute for Basic Standards, the Institute for Materials Research, the Institute for Applied Technology, the Institute for Computer Sciences and Technology, and the Office for Information Programs.

THE INSTITUTE FOR BASIC STANDARDS provides the central basis within the United States of a complete and consistent system of physical measurement; coordinates that system with measurement systems of other nations; and furnishes essential services leading to accurate and uniform physical measurements throughout the Nation's scientific community, industry, and commerce. The Institute consists of a Center for Radiation Research, an Office of Measurement Services and the following divisions:

Applied Mathematics — Electricity — Mechanics — Heat — Optical Physics — Nuclear Sciences² — Applied Radiation² — Quantum Electronics³ — Electromagnetics³ — Time and Frequency³ — Laboratory Astrophysics³ — Cryogenics³.

THE INSTITUTE FOR MATERIALS RESEARCH conducts materials research leading to improved methods of measurement, standards, and data on the properties of well-characterized materials needed by industry, commerce, educational institutions, and Government; provides advisory and research services to other Government agencies; and develops, produces, and distributes standard reference materials. The Institute consists of the Office of Standard Reference Materials and the following divisions:

Analytical Chemistry — Polymers — Metallurgy — Inorganic Materials — Reactor Radiation — Physical Chemistry.

THE INSTITUTE FOR APPLIED TECHNOLOGY provides technical services to promote the use of available technology and to facilitate technological innovation in industry and Government; cooperates with public and private organizations leading to the development of technological standards (including mandatory safety standards), codes and methods of test; and provides technical advice and services to Government agencies upon request. The Institute consists of a Center for Building Technology and the following divisions and offices:

Engineering and Product Standards — Weights and Measures — Invention and Innovation — Product Evaluation Technology — Electronic Technology — Technical Analysis — Measurement Engineering — Structures, Materials, and Life Safety⁴ — Building Environment⁴ — Technical Evaluation and Application⁴ — Fire Technology.

THE INSTITUTE FOR COMPUTER SCIENCES AND TECHNOLOGY conducts research and provides technical services designed to aid Government agencies in improving cost effectiveness in the conduct of their programs through the selection, acquisition, and effective utilization of automatic data processing equipment; and serves as the principal focus within the executive branch for the development of Federal standards for automatic data processing equipment, techniques, and computer languages. The Institute consists of the following divisions:

Computer Services — Systems and Software — Computer Systems Engineering — Information Technology.

THE OFFICE FOR INFORMATION PROGRAMS promotes optimum dissemination and accessibility of scientific information generated within NBS and other agencies of the Federal Government; promotes the development of the National Standard Reference Data System and a system of information analysis centers dealing with the broader aspects of the National Measurement System; provides appropriate services to ensure that the NBS staff has optimum accessibility to the scientific information of the world. The Office consists of the following organizational units:

Office of Standard Reference Data — Office of Information Activities — Office of Technical Publications — Library — Office of International Relations.

¹ Headquarters and Laboratories at Gaithersburg, Maryland, unless otherwise noted; mailing address Washington, D.C. 20234.

² Part of the Center for Radiation Research.

³ Located at Boulder, Colorado 80302.

⁴ Part of the Center for Building Technology.


15/3
acc.

Note on a Vibratory Phenomenon Arising in Transducer Calibration

Richard Kraft

Institute for Basic Standards
National Bureau of Standards
Washington, D.C. 20234

t. Technical note no 856



U.S. DEPARTMENT OF COMMERCE, Frederick B. Dent, *Secretary*
NATIONAL BUREAU OF STANDARDS, Richard W. Roberts, *Director*

Issued February 1975

Library of Congress Catalog Card Number: 74-600195

National Bureau of Standards Technical Note 856

Nat. Bur. Stand. (U.S.), Tech. Note 856, 13 pages (Feb. 1975)

CODEN: NBTNAE

U.S. GOVERNMENT PRINTING OFFICE
WASHINGTON: 1975

For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402
(Order by SD Catalog No. C13.46:856).

Price 65 cents

PREFACE

The work in this publication was conducted at the request of the Instrumentation Applications Section in support of the project "Space Shuttle Pogo Pressure Measuring System" for NASA George C. Marshall Space Flight Center.

Dr. H. S. Oser
Chief, Mathematical Analysis Section



Note on a Vibratory Phenomenon Arising in Transducer Calibration

Richard Kraft

By making appropriate physical approximations and idealizations a theoretical explanation is found for a vibratory phenomenon observed in calibrating pressure transducers inside thin liquid filled cylinders. The theoretical explanation requires proving the equivalence of two boundary initial problems which define the vibratory phenomenon. A short, general and complete proof of this equivalence is given.

Key words: boundary initial problems; measurement; pressure transducer; transducer calibration; vibration.

1. Introduction

The purpose of this note is to provide a theoretical explanation of an experiment connected with dynamic calibration of transducers. The dynamic calibration technique is concerned with small disturbances in a liquid filled metal tube that is open at the top and is mounted vertically on a vibration table, see figure 1b. Sound waves are generated in the liquid by vertical sinusoidal motion of the vibration table. The waves in the fluid are picked up by transducers attached along the side of the tube. It is readily observed that the pressures sensed by the transducer is a function of the applied acceleration, liquid density and position along the tube. Further experiments in which the tube was

capped at the top and completely filled with liquid, see figure 1a, were conducted in order to study the effect on the transducer of additional steady pressure. The data from this experiment shows (with reasonable experimental error) the interesting phenomenon that the pressure in an open tube of length, $L/2$, at any height above the vibration table is the same as that in a closed tube of length, L , at the same height above the vibration table when the closed and open tubes are vibrated at the same frequency. We emphasize that the induced vibrations are off resonance. In section 2, we give a mathematic formulation of these experiments which predicts this phenomenon. Since the mathematical description is based on a set of physical assumptions, see hypotheses I-III below. This prediction increases our confidence in these assumptions and provides a basis for interpretation and correlation of experimental data.

2. Analysis

We consider, for the sake of definiteness, the small disturbance motion imparted simultaneously to the fluid in both systems in figure 1 by a vibratory motion, $g(t)$, of the vibration table. Initially both systems are assumed to be at complete rest. The resulting small displacement field $\xi^C(x,t)$, that is setup inside the capped tube is precisely described, in the one dimensional non-dissipative case, using Lagrangian terminology [2, p. 479] by the equations

$$\partial_t^2 \xi^C(x,t) - c^2 \partial_x^2 \xi^C = 0 \quad (2.1i)$$

$$\text{for } 0 \leq x \leq L \quad \text{and} \quad t \geq 0 ,$$

$$\xi^C(x,0) = \partial_t \xi^C(x,0) = 0 \quad (2.1ii)$$

for $0 \leq x \leq L$,

$$\xi^C(0,t) = g(t) \quad (2.1iii)$$

for $t \geq 0$,

$$\xi^C(L,t) = g(t) \quad (2.1iv)$$

for $t \geq 0$.

Likewise, the precise description of the field, $\xi^0(x,t)$, inside the open tube is given by the equations

$$\partial_t^2 \xi^0(x,t) - c^2 \partial_x^2 \xi^0(x,t) = 0 \quad (2.2i)$$

for $x \leq L/2$ and $t \geq 0$,

$$\xi^0(x,0) = \partial_t \xi^0(x,0) = 0 \quad (2.2ii)$$

for $0 \leq x \leq L/2$,

$$\xi^0(0,t) = g(t) \quad (2.2iii)$$

for $t \geq 0$,

$$\partial_x \xi^0(L/2,t) = 0 \quad (2.2iv)$$

for $t \geq 0$.

In these equations $c^2 \equiv \left. \frac{dp}{d\rho} \right|_{\rho_0}$, where a zero subscript denotes an equilibrium value and $g(t)$ is sufficiently smooth to guarantee the existence of a twice continuously differentiable solution of the above equations.

The justification for the boundary conditions employed in the above equation derives from the following premises.

HYPOTHESIS I: The fluid at both ends of the closed tube and at the bottom of the open tube does not separate from the end of the tube. Hence the liquid at the capped ends of both tubes undergoes the same displacements as the end of the tube. This hypothesis can be rationalized by assuming "no-slip" conditions hold at the closed ends of the tube.

HYPOTHESIS II: The metal tubes are perfectly rigid, i.e. they do not vibrate, moreover they are rigidly attached to the vibrating table. In this situation the tube and vibrating table move together as a rigid body, with the same vibratory motion as the vibrating table. This is, in fact, approximately the case, because the wavelengths of sound in the tube is larger than in the liquid, and since the experiments were run below the fundamental resonance of the liquid itself, the entire length of the tube is only a fraction of a wavelength and therefore all parts of it vibrate approximately in phase.

Hypotheses I and II are the justification for boundary conditions (2.1iii,iv) and (2.2iii).

HYPOTHESIS III: The liquid at the open end of the uncapped tube behaves like a free surface, i.e. supports no stress.

This hypothesis implies

$$\partial_x \xi^0(L/2,t) = 0$$

at the open end of the uncapped tube.

Since (in Lagrangian terminology) a small disturbance pressure field, $p - p_0$, is related to its associated displacement field, $\xi(x,t)$, by the formula $p - p_0 = K \frac{\partial \xi}{\partial x}$, the experimental phenomenon described in section 1 is implied by the relationship

$$\xi^C(x,t) = \xi^0(x,t) \quad (2.3)$$

for $0 \leq x \leq L/2$ and for $t \geq 0$,

here $\xi^C(x,t)$, $\xi^0(x,t)$ are the fields determined by (2.1i-iv) and (2.2i-iv). That (2.3) is implied by the above equations can be seen in the following way.

First, a function $\xi^*(x,t)$ is defined in the whole interval $0 \leq x \leq L$, in terms of the twice continuously differentiable solution of (2.2i-iv) namely $\xi^0(x,t)$ by the reflection rule

$$\xi^*(x,t) \equiv \begin{cases} \xi^0(x,t) & \text{for } 0 \leq x \leq L/2 \\ \xi^0(L-x,t) & \text{for } L/2 \leq x \leq L. \end{cases} \quad (2.4)$$

It is readily verified using the properties of $\xi^0(x,t)$ that ξ^* is a twice continuously differentiable function in the interval $0 \leq x \leq L$. Also, since (2.4) implies

$$\partial_t^2 \xi^*(x,t) = \begin{cases} \partial_t^2 \xi^0(x,t) & \text{for } 0 \leq x \leq L/2 \\ \partial_t^2 \xi^0(L-x,t) & \text{for } L/2 \leq x \leq L \end{cases}$$

and

$$\partial_x^2 \xi^*(x,t) = \begin{cases} \partial_x^2 \xi^0(x,t) & \text{for } 0 \leq x \leq L/2 \\ \partial_x^2 \xi^0(L-x,t) & \text{for } L/2 \leq x \leq L, \end{cases}$$

equation (2.2i) implies $\xi^*(x,t)$ satisfies (2.1i) in the interval $0 \leq x \leq L$. By using (2.4) to evaluate the quantities $\xi^*(0,t)$ and $\xi^*(L,t)$ it is seen that $\xi^*(x,t)$ satisfies the boundary conditions (2.1iii) and (2.1iv). Finally, in the same way (2.4) implies $\xi^*(x,0)$ satisfies the initial condition (2.1ii). Since it has now been shown that $\xi^*(x,t)$ is twice continuously differentiable and satisfies (2.1i-iv), it must be the unique solution of these equations. Hence $\xi^*(x,t) = \xi^C(x,t)$ and (2.3) follows from this by employing (2.4) again.

We note that the proof of (2.3) goes through in the same way when general initial conditions are substituted into (2.1ii, 2.2ii) and even if a Rayleigh type dissipative term i.e. $-k^2 \partial_t \xi$ is inserted in the wave equations.

In conclusion we remark that the boundary initial problems that determine $\xi^C(x,t)$ and $\xi^0(x,t)$ can be directly solved "by the variation of parameter method" and (2.3) is seen to be true by this method also. However, this direct method of establishing (2.3) is considerably longer and more complicated than the proof above.

3. Discussion

The key part in the previous derivation of relation (2.3) was played by the boundary condition (2.2iv) as this was the condition that suggested and permitted the reflection type definition of ξ^* given in (2.4). The reflection rule leads one to consider the extent to which the relation $\xi^C(0,t) = \xi^C(L,t)$ holds and as stated in hypothesis II there is good reason to suppose it is a good approximation. With a knowledge of this relation and (2.4) the completion of the derivation of (2.3) is

straightforward. The physical meaning of the condition (2.2iv) is that the fluid at the top of the open tube is completely free to move and thus no stress, which is proportional to $\frac{\partial \xi^0}{\partial x}$, can be induced in it. Thus, although $\frac{\partial \xi^0}{\partial x}$ is proportional to the fluid pressure, it is misleading to assert that $\frac{\partial \xi^0}{\partial x} = 0$ is true because the atmosphere is in some way forcing the fluid at the fluid-air interface to remain in equilibrium. Indeed, the atmosphere has no appreciable effect on the fluid and that is why it behaves as a free surface.

Experiments aimed at measuring the speed of sound in fluids have been made in fluid filled cylinders [1 and earlier citations therein] and it has been observed that the elasticity of the tube wall effects these measurements. The above analysis shows that the phenomenon described in this note will be disturbed to the extent that such elastic effects are present.

In closing it is noted that the Lagrangian approach used in section 2 is advantageous because the boundary conditions are most naturally described in terms of displacements, and displacement is the chief dependent variable in the Lagrangian formalism.

Acknowledgement

The author is grateful to Martin Greenspan and Paul S. Lederer for passing on their experience and insights related to the above problem.

References

1. Gronwall, T. H.; The Longitudinal Vibrations of a Liquid Contained in a Tube with Elastic Walls; Phys. Rev.; Vol. 30; July 1927.
2. Lamb, H.; Hydrodynamics; Dover Publications; 1945.




Figure 1a, b: Vibrating cylinders

U.S. DEPT. OF COMM. BIBLIOGRAPHIC DATA SHEET		1. PUBLICATION OR REPORT NO. NBS TN 856	2. Gov't Accession No.	3. Recipient's Accession No.
4. TITLE AND SUBTITLE Note on a Vibratory Phenomenon Arising in Transducer Calibration			5. Publication Date February 1975	
			6. Performing Organization Code	
7. AUTHOR(S) Richard Kraft			8. Performing Organ. Report No.	
9. PERFORMING ORGANIZATION NAME AND ADDRESS NATIONAL BUREAU OF STANDARDS DEPARTMENT OF COMMERCE WASHINGTON, D.C. 20234			10. Project/Task/Work Unit No. 2050141	
			11. Contract/Grant No. 2050141	
12. Sponsoring Organization Name and Complete Address (Street, City, State, ZIP) National Bureau of Standards Department of Commerce Washington, D. C. 20234			13. Type of Report & Period Covered Final	
			14. Sponsoring Agency Code	
15. SUPPLEMENTARY NOTES Library of Congress Catalog Card Number: 74-600195				
16. ABSTRACT (A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here.) By making appropriate physical approximations and idealizations a theoretical explanation is found for a vibratory phenomenon observed in calibrating pressure transducers inside thin liquid filled cylinders. The theoretical explanation requires proving the equivalence of two boundary initial problems which define the vibratory phenomenon. A short, general and complete proof of this equivalence is given.				
17. KEY WORDS (six to twelve entries; alphabetical order; capitalize only the first letter of the first key word unless a proper name; separated by semicolons) Boundary initial problems; measurement; pressure transducer; transducer calibration; vibration.				
18. AVAILABILITY <input checked="" type="checkbox"/> Unlimited <input type="checkbox"/> For Official Distribution. Do Not Release to NTIS <input type="checkbox"/> Order From Sup. of Doc., U.S. Government Printing Office Washington, D.C. 20402, SD Cat. No. C13, 46:856 <input type="checkbox"/> Order From National Technical Information Service (NTIS) Springfield, Virginia 22151			19. SECURITY CLASS (THIS REPORT) UNCLASSIFIED	
			21. NO. OF PAGES 13	
			20. SECURITY CLASS (THIS PAGE) UNCLASSIFIED	
			22. Price 65 cents	

NBS TECHNICAL PUBLICATIONS

PERIODICALS

JOURNAL OF RESEARCH reports National Bureau of Standards research and development in physics, mathematics, and chemistry. Comprehensive scientific papers give complete details of the work, including laboratory data, experimental procedures, and theoretical and mathematical analyses. Illustrated with photographs, drawings, and charts. Includes listings of other NBS papers as issued.

Published in two sections, available separately:

• **Physics and Chemistry (Section A)**

Papers of interest primarily to scientists working in these fields. This section covers a broad range of physical and chemical research, with major emphasis on standards of physical measurement, fundamental constants, and properties of matter. Issued six times a year. Annual subscription: Domestic, \$17.00; Foreign, \$21.25.

• **Mathematical Sciences (Section B)**

Studies and compilations designed mainly for the mathematician and theoretical physicist. Topics in mathematical statistics, theory of experiment design, numerical analysis, theoretical physics and chemistry, logical design and programming of computers and computer systems. Short numerical tables. Issued quarterly. Annual subscription: Domestic, \$9.00; Foreign, \$11.25.

DIMENSIONS/NBS (formerly Technical News Bulletin)—This monthly magazine is published to inform scientists, engineers, businessmen, industry, teachers, students, and consumers of the latest advances in science and technology, with primary emphasis on the work at NBS.

DIMENSIONS/NBS highlights and reviews such issues as energy research, fire protection, building technology, metric conversion, pollution abatement, health and safety, and consumer product performance. In addition, **DIMENSIONS/NBS** reports the results of Bureau programs in measurement standards and techniques, properties of matter and materials, engineering standards and services, instrumentation, and automatic data processing.

NONPERIODICALS

Monographs—Major contributions to the technical literature on various subjects related to the Bureau's scientific and technical activities.

Handbooks—Recommended codes of engineering and industrial practice (including safety codes) developed in cooperation with interested industries, professional organizations, and regulatory bodies.

Special Publications—Include proceedings of high-level national and international conferences sponsored by NBS, precision measurement and calibration volumes, NBS annual reports, and other special publications appropriate to this grouping such as wall charts and bibliographies.

Applied Mathematics Series—Mathematical tables, manuals, and studies of special interest to physicists, engineers, chemists, biologists, mathematicians, computer programmers, and others engaged in scientific and technical work.

National Standard Reference Data Series—Provides quantitative data on the physical and chemical properties of materials, compiled from the world's literature and critically evaluated. Developed under a world-wide program coordinated by NBS. Program under authority of National Standard Data Act (Public Law 90-396).

Building Science Series—Disseminates technical information developed at the Bureau on building materials, components, systems, and whole structures. The series presents research results, test methods, and performance criteria related to the structural and environmental functions and the durability and safety characteristics of building elements and systems.

Technical Notes—Studies or reports which are complete in themselves but restrictive in their treatment of a subject. Analogous to monographs but not so comprehensive in scope or definitive in treatment of the subject area. Often serve as a vehicle for final reports of work performed at NBS under the sponsorship of other government agencies.

Voluntary Product Standards—Developed under procedures published by the Department of Commerce in Part 10, Title 15, of the Code of Federal Regulations. The purpose of the standards is to establish nationally recognized requirements for products, and to provide all concerned interests with a basis for common understanding of the characteristics of the products. The National Bureau of Standards administers the Voluntary Product Standards program as a supplement to the activities of the private sector standardizing organizations.

Federal Information Processing Standards Publications (FIPS PUBS)—Publications in this series collectively constitute the Federal Information Processing Standards Register. The purpose of the Register is to serve as the official source of information in the Federal Government regarding standards issued by NBS pursuant to the Federal Property and Administrative Services Act of 1949 as amended, Public Law 89-306 (79 Stat. 1127), and as implemented by Executive Order 11717 (38 FR 12315, dated May 11, 1973) and Part 6 of Title 15 CFR (Code of Federal Regulations). FIPS PUBS will include approved Federal information processing standards information of general interest, and a complete index of relevant standards publications.

Consumer Information Series—Practical information, based on NBS research and experience, covering areas of interest to the consumer. Easily understandable language and illustrations provide useful background knowledge for shopping in today's technological marketplace.

NBS Interagency Reports—A special series of interim or final reports on work performed by NBS for outside sponsors (both government and non-government). In general, initial distribution is handled by the sponsor; public distribution is by the National Technical Information Service (Springfield, Va. 22151) in paper copy or microfiche form.

Order NBS publications (except Bibliographic Subscription Services) from: Superintendent of Documents, Government Printing Office, Washington, D.C. 20402.

BIBLIOGRAPHIC SUBSCRIPTION SERVICES

The following current-awareness and literature-survey bibliographies are issued periodically by the Bureau:

Cryogenic Data Center Current Awareness Service (Publications and Reports of Interest in Cryogenics). A literature survey issued weekly. Annual subscription: Domestic, \$20.00; foreign, \$25.00.

Liquefied Natural Gas. A literature survey issued quarterly. Annual subscription: \$20.00.

Superconducting Devices and Materials. A literature survey issued quarterly. Annual subscription: \$20.00. Send subscription orders and remittances for the pre-

ceding bibliographic services to the U.S. Department of Commerce, National Technical Information Service, Springfield, Va. 22151.

Electromagnetic Metrology Current Awareness Service (Abstracts of Selected Articles on Measurement Techniques and Standards of Electromagnetic Quantities from D-C to Millimeter-Wave Frequencies). Issued monthly. Annual subscription: \$100.00 (Special rates for multi-subscriptions). Send subscription order and remittance to the Electromagnetic Metrology Information Center, Electromagnetics Division, National Bureau of Standards, Boulder, Colo. 80302.

U.S. DEPARTMENT OF COMMERCE
National Bureau of Standards
Washington, D.C. 20234

OFFICIAL BUSINESS

Penalty for Private Use, \$300

POSTAGE AND FEES PAID
U.S. DEPARTMENT OF COMMERCE
COM-215

Third Class Mail

